



NovitaTech Engineering

1 300 855 585, Direct Line: (08) 8243 8262, fax (08) 8243 8208

Phil Mundy
Product Design Group Inc.
102-366 E Kent Ave S
Vancouver, British Columbia
Canada V5X 4N6

04th March 2005

Dear Phil

Reference: testing the PDG Eclipse 600, Stellar and Bentley manual wheelchairs, our laboratory reference are 490869, 490870 and 490871 respectively.

This letter is to advise you that this Laboratory has finished the testing of your products using the Standard methods of AS3696-1:1990, AS 3696-3:1990, AS 3696-5:1989, AS 3696-8:1998 and ISO 7176-7:1996.

Three official twelve-page (12) reports are attached, identifying the samples, the test documents and the testing facility. All conformances with the Australian standards are indicated in the report.

The PDG Eclipse 600 manual wheelchair as represented by the test sample and incorporating all modifications to its labelling and documentation complies with requirements of AS3695:1992 *"Wheelchairs – Product Requirements"*.

The PDG Stellar manual wheelchair as represented by the test sample and incorporating all modifications to its labelling and documentation complies with requirements of AS3695:1992 *"Wheelchairs – Product Requirements"*.

The PDG Bentley manual wheelchair as represented by the test sample and incorporating all modifications to its labelling and documentation complies with requirements of AS3695:1992 *"Wheelchairs – Product Requirements"*.

Please do not hesitate to contact us if you have any further enquiries or comments regarding the test.

Kind regards



Quality
Endorsed
Company
ISO 9001

Alex Svistounov
ME (Mech), M.I.E. Aust.
Senior Testing Engineer
NATA signatory

Location: 171 Days Road, Regency Park SA 5010 **Postal:** PO Box 2438, Regency Park SA 5942

Phone: (08) 8243-8262 (office), **Fax:** (08) 8243-8208

Web site: <http://www.novitatech.org.au/> **Email:** testing@novita.org.au