



Torr Brown  
Product Design Group  
Unit 103-318 East Kent Ave. South  
Vancouver, BC, Canada V5X4N6

21<sup>st</sup> May 2014

Dear Torr,

Reference: review of testing on the Product Design Group Eclipse manual wheelchair (273 kg / 600 lb), our Laboratory reference is 492108-2.

This letter is to advise you that this Laboratory has reviewed the testing of your product to the following Standard methods: ISO7176-1:1999, 7176-3:2003, 7176-5:2008, 7176-7:1998, 7176-8:1998, 7176-13:1989, 7176-15:1996 & 7176-22:2000 (except materials and surface finishes).

This review is based on the original test report # 492108 dated 21<sup>st</sup> May 2014 by this laboratory for testing performed to AS3695.1:2011 test standards.

The official twelve (12) page report, dated 21<sup>st</sup> May 2014 and referencing job number 492108-2, identifies the sample, the test documents and the testing facility. All conformances with the ISO standards are indicated in the report.

The Product Design Group Eclipse manual wheelchair as represented by the test sample, including the supplied documentation and labelling as stated in the test report, complies with the relevant requirements of the above referenced standards (excluding the methods indicated in the report as "not tested/ not measured") for user mass 273kg / 600lb.

Please do not hesitate to contact us if you have any enquiries or comments regarding the test.

Kind regards,

Wayne Wurfel  
Senior Test Technician, Test Laboratory  
Authorised signatory

**NovitaTech (a division of Novita Children's Services)**  
171 Days Road, Regency Park, South Australia 5010  
PO Box 2438, Regency Park, South Australia 5942  
Telephone 1300 855 585 • Fax (08) 8243 8208  
enquiries@novita.org.au • www.novitatech.org.au  
ABN 53 137 537 636

