



Torr Brown  
Product Design Group  
Unit 103-318 East Kent Ave. South  
Vancouver, BC, Canada V5X4N6

17<sup>th</sup> November 2014

Dear Torr,

Reference: review of testing on the Product Design Group Fuze T50 manual wheelchair (204 kg / 450 lb), our Laboratory reference is 491806-1.

This letter is to advise you that this Laboratory has reviewed the testing of your product to the following Standard methods: ISO7176-1:1999, 7176-3:2003, 7176-5:2008, 7176-7:1998, 7176-8:1998, 7176-13:1989, 7176-15:1996 & 7176-22:2000 (except ignitibility, and materials).

This review is based on the original test report # 491430 dated 15<sup>th</sup> December 2010 by this laboratory for testing performed to AS3695:1992 test standards.

The official thirteen (13) page report, dated 7<sup>th</sup> November 2014 and referencing job number 491806-1, identifies the sample, the test documents and the testing facility. All conformances with the ISO standards are indicated in the report.

The Product Design Group Fuze T50 manual wheelchair as represented by the test sample, including the supplied documentation and labelling as stated in the test report, complies with the relevant requirements of the above referenced standards (excluding the methods indicated in the report as "not tested/ not measured") for user mass 204 kg / 450 lb.

Please do not hesitate to contact us if you have any enquiries or comments regarding the test.

Kind regards,

Wayne Wurfel  
Senior Test Technician, Test Laboratory  
Authorised signatory



Accreditation No. 2953

**NovitaTech (a division of Novita Children's Services)**  
171 Days Road, Regency Park, South Australia 5010  
**PO Box 2438, Regency Park, South Australia 5942**  
Telephone 1300 855 585 • Fax (08) 8243 8208  
enquiries@novita.org.au • www.novitatech.org.au  
ABN 53 137 537 636

