



Torr Brown, Chief Engineer
PDG Product Design Group Inc.
103 – 318 East Kent Avenue South,
Vancouver, BC Canada
V5X 4N6

30th June 2016

Dear Torr,

Reference: Testing the PDG Product Design Group Inc. Stellar GL (G2 version) manual wheelchair to 114 kg (250 lbs) user weight, our Laboratory reference is 492559.

This letter is to advise you that this Laboratory has finished testing your product to AS/NZS 3695.1:2011 Wheelchairs part 1: Requirements and test methods for manual wheelchairs: including references to the following standards:

AS/NZS ISO 7176 Sections 1, 3, 8, 11, 16, 22 & 26,
AS/NZS 3696 Section 19,
AS3696 Sections 5 & 13
ISO 7176 Section 7,
for 114 kg / 250 lbs user weight.

The PDG Product Design Group Inc. Stellar GL (G2 version) manual wheelchair as represented by the test sample, including the supplied documentation and labelling as stated in the test report, complies with the above referenced standards conditional that the following requirements are met:

This wheelchair is available with a transit option (Transit Tie Down System, TTDS). The owners' manual states that this wheelchair has been dynamically tested in a forward facing mode for 30 mph frontal impact test. A separate report for compliance to this testing should be made available to the purchaser upon request.

If the transit option is fitted to the wheelchair, and the approved anchorage points are identified with tie-down anchorage point symbols (Wheelchair Tie-down and Occupant Restraint Systems WTORS symbols) no further labelling is required to meet AS/NZS standards compliance.

If the transit option is not fitted, the wheelchair must be labelled with the following warning:

“Not AS/NZS 3696.19 Compliant”

The label shall conform to the requirements of Clauses 8.2(d)(ii), 8.2 (d)(ii) A,B,C and 8.2 (e) (i, ii, iii) as below:

- Clauses, 8.2 (d) (ii) For AS/NZS 3696.19 Non-compliant chairs, warning label with the words 'Not AS/NZS 3696.19 compliant'
- Clause 8.2 (d)(ii) A,B,C Label to have correct font size, contrast & location on chair
- Clause 8.2 (e) (i, ii, iii) Label/s to be permanent and durable

NovitaTech (a division of Novita Children's Services)
171 Days Road, Regency Park, South Australia 5010
PO Box 2438, Regency Park, South Australia 5942
Telephone 1300 855 585 • Fax (08) 8243 8208
enquiries@novita.org.au • www.novitatech.org.au
ABN 53 137 537 636





To fully comply with local requirements of AS/NZS 3695.1:2011, the following requirements need to be met if selling the product into AS/NZ markets:

- User instructions should include a warning regarding risk of increased surface temperatures if the wheelchair is exposed to direct sunlight.
- User instructions should include information on the recycling of the wheelchair.

A copy of the test report/s must be made available to the purchaser if requested.

Please do not hesitate to contact us if you have any enquiries or comments regarding this testing.

Kind regards,

Wayne Wurfel

Senior Test Technician, Authorised signatory

NovitaTech (a division of Novita Children's Services)
171 Days Road, Regency Park, South Australia 5010
PO Box 2438, Regency Park, South Australia 5942
Telephone 1300 855 585 • Fax (08) 8243 8208
enquiries@novita.org.au • www.novitatech.org.au
ABN 53 137 537 636

